

adrienne


From: Alice Lara [alice@sads.org]
Sent: Monday, March 15, 2010 10:46 AM
To: adrienne@sads.org
Subject: ICD Announcement from Boston Scientific

Having trouble viewing this email? [Click here](#)

A Message from The SADS Foundation

Dear Alice,

Last Friday in our e-newsletter, we sent important information from Boston Scientific regarding ICD's. In the e-mail, the link to the announcement was incorrect and thus some people may not have been able to access it. We want to be sure you were able read the announcement, so we have sent it once more.



[ICD Advisory from Boston Scientific:](#)

Boston Scientific recently contacted SADS to let us know that ICD's from the COGNIS® and TELIGEN® families that are implanted under the chest muscle (rare) may have some weakening over time with the possibility of delivering inappropriate shocks. Click [here](#) to read the full statement from Boston Scientific.

If you have any questions, please e-mail Adrienne at adrienne@sads.org or call 1-800-786-7723.

Sincerely,

A handwritten signature in cursive script that reads "Alice Lara".

Alice Lara
SADS Foundation

[Forward email](#)

SafeUnsubscribe®

This email was sent to alice@sads.org by alice@sads.org.
Instant removal with [SafeUnsubscribe™](#) | [Privacy Policy](#).

Email Marketing by



SADS Foundation | 508 E. South Temple, Suite 20 | Salt Lake City | UT | 84102

THIS IS A TEST EMAIL ONLY.

This email was sent by the author for the sole purpose of testing a draft message. If you believe you have received the message in error, please contact the author by replying to this message. Constant Contact takes reports of abuse very seriously. If you wish to report abuse, please forward this message to abuse@constantcontact.com.